



NATIONAL TYPE EVALUATION PROGRAM

Certificate of Conformance

for Weighing and Measuring Devices

For:

Non-Computing Scale
Counter Bench
Models: SW-5, SW-10, SW-20, SW-50 & SW-RS
 n_{max} : 2 500
 e_{min} : 0.002 lb / 0.001 kg / 1 g / 0.05 oz
Capacity: 5 lb to 50 lb* / 20 kg / 20 000 g / 800 oz
Platform: 7 1/4 in x 9 1/4 in
Accuracy Class: III

Submitted By:

CAS USA, Corporation
99-A Murray Hill Parkway
East Rutherford, NJ 07073
Tel: 201-933-9002
Fax: 201-933-9025
Contact: William Moutenot
Email: bill@cas-usa.com
Web site: www.cas-usa.com

Standard Features and Options**Standard Features:**

- Semi-automatic (push-button) Zero
- Automatic Zero Setting Mechanism (AZSM)
- Initial Zero Setting Mechanism
- Semi-automatic (push-button) Tare
- Battery Saving Feature (auto shut off)
- AC/DC Adapter
- Customer Display
- lb, oz, kg, g Capability
- Liquid Crystal Display (LCD)
- RS232C interface communication (Identified with an -RS suffix)

***Specific Models and Capacities:**

Model	Capacity
SW-5	5 x 0.002 lb / 2 x 0.001 kg / 2000g x 1 g / 80 x 0.05 oz
SW-10	10 x 0.005lb / 5 x 0.002 kg / 5000g x 2 g / 160 x 0.1 oz
SW-20	20 x 0.01lb / 10 x 0.005 kg / 10000g x 5 g / 400 x 0.2 oz
SW-50	50 x 0.02 lb / 20 x 0.01 kg / 20000g x 10 g / 800 x 0.5 oz

Load Cells Used:

- CAS Models SW-2L, SW-5L, SW-10L and SW-20L

Temperature Range: -10 °C to 40 °C (14 °F to 104 °F)

This device was evaluated under the National Type Evaluation Program and was found to comply with the applicable technical requirements of "NIST Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices." Evaluation results and device characteristics necessary for inspection and use in commerce are on the following pages.

Jerry Buendel
Chairman, NCWM, Inc.

Ronald Hayes
Chairman, National Type Evaluation Program Committee
Issued: May 6, 2016

1135 M Street, Suite 110 / Lincoln, Nebraska 68508

The National Conference on Weights and Measures (NCWM) does not approve, recommend or endorse any proprietary product or material, either as a single item or as a class or group. Results shall not be used in advertising or sales promotion to indicate explicit or implicit endorsement of the product or material by the NCWM.



CAS USA, Corporation

Non-Computing Scale / SW-5, SW-10, SW-20, SW-50 & SW-RS

Application: General purpose bench and counter scale.

Identification: The required information is located on a metal badge riveted to the side of the scale and on the face plate adjacent to the display.

Sealing: A metal plate covers the calibration switch on the bottom of the scale. A screw with a hole drilled across the end is placed through the scale from the top with the end protruding through the metal plate. A wire seal can be passed through the hole on the end of the screw and through a plastic tab that also protrudes through the metal plate to secure access to the calibration switch.

Test Conditions: This Certificate supersedes Certificate of Conformance 99-002A1 and is issued to recognize a change in the model designation when the device is equipped with an RS232C interface port. There was no change to the scale, so no testing was deemed necessary. Previous test conditions are listed below for reference.

Certificate of Conformance Number 99-002A1: This certificate supersedes NTEP Certificate of Conformance Number 99-002 and is issued to add RS-232C interface communication to the existing Certificate. The emphasis of the evaluation was on performance testing of the RS-232C interfacing to (07-002) PC America Point of Sale System.

Certificate of Conformance 99-002: Two Model SW scales were submitted for evaluation (5 lb x 0.002 lb and a 50 lb x 0.02 lb). The emphasis of the evaluation was on device design, operation, compliance with marking requirements, and influence factor requirements. Several Increasing/decreasing load and shift tests were conducted. The scales were tested over a range of temperature from -10 °C to 40 °C (14 °F to 104 °F). A load of approximately one-half capacity was applied to the scales over 100 000 times. The scales were tested periodically over this time. Additionally, tests were conducted using 100 VAC and 130 VAC power supplies.

Evaluated By: E. Matthews (OH), W. West (OH) 99-002; M. Kelley (OH) 99-002A1

Type Evaluation Criteria Used: NIST, Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices, 2016. NCWM, Publication 14: Weighing Devices, 2016.

Conclusion: The results of the evaluation and information provided by the manufacturer indicate the device complies with applicable requirements.

Information Reviewed By: J. Truex (NCWM) 99-002A1, 99-002A2

Example of Device:



SW-Series